

Outdoor Pyranometer Calibration (OPC) Service

February 2024




Member of Group



Introduction- Outdoor Pyranometer Calibration

- ❑ Calibrating pyranometers in real-world conditions helps account for environmental factors that can affect their accuracy, such as temperature, humidity, and atmospheric conditions.
- ❑ Outdoor Pyranometer Calibration (OPC) helps to shorten the time to delivering sensors to lab.
- ❑ This work is essential in the solar energy industry to maintain the precision of solar irradiance measurements, which are critical for system performance assessment, energy forecasting, and research purposes.



Outdoor Pyranometer Calibration (OPC)

ExTEL
In-house

Traceable & calibrated
Reference
Pyranometer



Indoor
Verification



ExTEL
On-Site
Service
(OPC)

Outdoor
Calibration

ISO 9847


Calibrated & Traceable Certificate



- Trace to PMOD/WRC
- Refer to Secondary Standard Pyranometer (CMP22)
- Follow ISO 9847:1992(E) calibration procedure



- Class AAA Sun Simulator
- Follow IEC 60891 standard
- Uncertainty $< 0.9\%$ (0.07 uV/W/m²)



Traceable & calibrated Reference
Pyranometer (CMP22)

Calibration Targets:

- ① Other Standard Pyranometer
- ② Reference Radiation crystal silicon cell



OPC - ISO 9847 Outdoor Calibration Standard

CMP22 Traceable and Calibrated

KIPP & ZONEN
CALIBRATION CERTIFICATE

CERTIFICATE NUMBER	019/PS/00165
PYRANOMETER MODEL	CMP22
SERIAL NUMBER	100865
SENSITIVITY	6.72 $\mu\text{W}/\text{m}^2$ at normal incidence on horizontal pyranometer
IMPEDANCE	30 Ω
REFERENCE PYRANOMETER	Kipp & Zonen CM 22 in ISO active from 01 January 2017
CALIBRATION DATE	05 February 2018
CLASSIFICATION	ISO 9002, Secondary standard

Calibration procedure:
The calibration was performed in a dark room by direct comparison with a reference pyranometer under an artificial sun for an AC voltage of 230 V. The reference pyranometer is a Kipp & Zonen CM 22 in ISO active from 01 January 2017. The reference pyranometer is a Kipp & Zonen CM 22 in ISO active from 01 January 2017. The reference pyranometer is a Kipp & Zonen CM 22 in ISO active from 01 January 2017.

Hierarchy of traceability:
The reference pyranometer was compared with the reference pyranometer under an artificial sun for an AC voltage of 230 V. The reference pyranometer is a Kipp & Zonen CM 22 in ISO active from 01 January 2017. The reference pyranometer is a Kipp & Zonen CM 22 in ISO active from 01 January 2017.

Justification of total instrument calibration uncertainty:
The calibration uncertainty is the result of the calibration procedure. The calibration uncertainty is the result of the calibration procedure. The calibration uncertainty is the result of the calibration procedure.

Notes:
The calibration certificate is valid for the calibration of the reference pyranometer. The calibration certificate is valid for the calibration of the reference pyranometer. The calibration certificate is valid for the calibration of the reference pyranometer.

Delft, The Netherlands, 05 February 2018

Jeroen Mas
In charge of calibration facility

V. Tromp
On charge of test


Kipp & Zonen B.V.
Dordrechtse wijk, 2618 XH Delft
The Netherlands
+31 15 2755 210
www.kippzonen.com

MTR no. 18.0005.00017.01
Team Manager
Number of MTR

OPC field calibration – high throughput; 8+ units/day

ISO 9847

Allows for calibration of a pyranometer indoors (as at the factory in Delft) against a reference pyranometer of similar type, or outdoors. Outdoors a field pyranometer can be calibrated over several days against a 'reference' pyranometer of similar or (ideally) higher quality with a reliable, recent, and traceable calibration.



OPC Procedure [1/5]

Disassembly/ Disconnect

- Photo, S/N & Sensitivity record
- Remove pyranometer

Handled by
Customer local team

Record



Removal



Sensitivity

Serial Number (S/N)




OPC Procedure [2/5]

Calibration Preparation

- Appearance Check
- Basic clean (Desiccant prepare by Customer)
- Setup next to CMP22 leveling

Handled by ExTEL Energy Engineers

Setup on plate



Leveling




Bubble is in center



OPC Procedure [5/5]

- ① Each individual **Pyranometer** is correlated to Reference **CMP22** from real-time on-site outdoor measurement



- ② Confirm **Zero Point** reading if close to 0 W/m² (within +/-3W/m²)

Dark Room Condition



- ③ Calculate **Correction Factor** for each individual CMP11

If $|Correlation Ratio - 1| \leq 1.0\%$
No correction needed

If $|Correlation Ratio - 1| > 1.0\%$
Suggest to implement
Correction Factor to SCADA

Corrected Irradiance [W/m²]
= Old Irradiance reading
(SCADA) x **Correction Factor**

For Example -

- ① Analyzing correlation ratio to CMP22 reference → 102.1%
- ② Confirming Zero Point reading → -1.0W/m² (within +/-3W/m²)
- ③ $|Correlation Ratio 102.1\% - 1| > 1.0\%$, **Correction Factor = 1/102.1% = 0.979**






Notice

This material contains confidential information. You may not copy or disclose to any third party without the prior written consent of ExTEL Energy Co. Ltd.



Confidential



Member of Group

